

Certificate of Accreditation

International Accreditation Japan (IAJapan) hereby accredits the following conformity assessment body as a Reference Material Producer of Japan Calibration Service System.

Accreditation Identification: JCSS 0297 RMP

Name of Conformity Assessment Body: Calibration Laboratory, NIPPON GREASE, Co., Ltd.

Name of Legal Entity: NIPPON GREASE, Co., Ltd.

Location of Conformity Assessment Body: 1-12-4 Suehiro-cho, Tsurumi-ku, Yokohama-shi,

Kanagawa 230-0045, JAPAN

Scope of Accreditation: Viscosity (as the following pages)

Accreditation Requirement: ISO 17034:2016*

* The relevant accreditation requirements described in the Accreditation

Scheme Document for JCSS-RMP are also applied.

Effective Date of Accreditation: 2023-06-17

Expiry Date of Accreditation: 2027-06-16

Date of Initial Accreditation: 2012-12-25

L. Saile

SAITO Kazunori

Chief Executive, International Accreditation Japan (IAJapan)
National Institute of Technology and Evaluation

⁻ International Accreditation Japan (IAJapan) is a laboratory accreditation body which has signed MRAs of ILAC (International Laboratory Accreditation Cooperation) and APAC (Asia Pacific Accreditation Cooperation).

⁻ MRA requirements are, in addition to relevant international standards and guides, requirements for participation in proficiency testing programs, surveillance and reassessment, and the policy for the traceability of measurement for MRA purpose.

⁻ This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system in accordance with the recognized International Standard ISO 17034:2016.

⁻ The latest accreditation information is publicly available on IAJapan Website as an accreditation certificate.

<Reference Material Producer>

Category: Viscosity

Date of Initial Accreditation of the Field: 2012-12-25

Type: Certified Reference Material

The Approach Used to Assign a Property Value: Measurement by a Single Method in a Single Laboratory (ISO 17034:2016

7.12.3 NOTE 1 d)

Sub- Category	Property	Range	4	Expanded Uncertainty (Level of Confidence Approximately 95 %)	Characterizatio Techniques
Viscosity Standards	Standard Liquids for Calibrating Viscometer	From 0.5 mm ² /s	From 20°C	0.17.0/	Capillary viscometer
		less than 1.8 mm ² /s	up to 40°C	0.17 %	
		From 1.8 mm ² /s	From 20°C	0.10.0/	
		less than 5.0 mm ² /s	up to 40°C	0.19 %	
		From 5.0 mm ² /s	From 20°C	0.14 %	
		less than 20 mm ² /s	up to 40°C		
		From 20 mm ² /s	From 20°C	0.19 % 0.20 %	
		less than 150 mm ² /s	up to 40°C		
		From 150 mm ² /s	From 20°C		
		less than 940 mm ² /s	up to 40°C		
		From 940 mm ² /s	From 20°C	0.24 %	
		less than 2 000 mm ² /s	up to 40°C		
		From 2 000 mm ² /s	From 20°C	0.39 %	
		less than 14 000 mm ² /s	up to 40°C		
		From 14 000 mm ² /s	From 20°C	1.15 %	
		less than 52 000 mm ² /s	up to 40°C		
		From 52 000 mm ² /s	From 20°C	0.69 %	
		up to 192 000 mm ² /s	up to 40°C		
	Standard Liquids	From 0.4 mPa·s	From 20°C	0.17.0/	Capillary viscometer
	for Calibrating Viscometer	less than 1.4 mPa·s	up to 40°C	0.17 %	
		From 1.4 mPa·s	From 20°C	0.19 %	
		less than 4.1 mPa·s	up to 40°C		
		From 4.1 mPa·s	From 20°C		
		less than 17 mPa*s	up to 40°C	0.15 %	
	-				
		From 17 mPa·s	From 20°C	0.19 %	
		less than 130 mPa·s	up to 40°C		
		From 130 mPa·s	From 20°C	0.20 % 0.25 % 0.39 %	
		less than 820 mPa·s	up to 40°C		
	#C	From 820 mPa·s	From 20°C		
		less than 1 800 mPa·s	up to 40°C		
		From 1 800 mPa·s	From 20°C		
		less than 12 000 mPa·s	up to 40°C		
		From 12 000 mPa·s			
			From 20°C	1.16 %	
		less than 46 000 mPa·s	up to 40°C	MINIOTAL BOARD AND AND AND AND AND AND AND AND AND AN	
		From 46 000 mPa·s	From 20°C	0.69 %	
		up to 168 000 mPa·s	up to 40°C	0.05 70	